

UK Declaration of Conformity



We declare under our sole responsibility that the products:

AMI CACE	AMI pH/mV:pH/mV	AMI-II CACE
AMI Codes-II	AMI pH-Redox	AMI-II CACE DG
AMI Codes-II CC	AMI Powercon	
AMI Codes-II O₃	AMI Rescon	AMI INSPECTOR Conductivity
AMI Codes-II TC	AMI SAC254	AMI INSPECTOR Hydrogen
AMI Deltacon DG	AMI Sequencer	AMI INSPECTOR Oxygen
AMI Deltacon Power	AMI Silica	AMI INSPECTOR pH
AMI Hydrazine	AMI Silitrace	AMI INSPECTOR Pharmacon
AMI Hydrogen QED	AMI Silitrace Ultra	AMI INSPECTOR Resistivity
AMI Inducon	AMI Soditrace	AMU Inducon
AMI ISE Universal	AMI Sodium A	AMU-II Oxytrace
AMI LineTOC / Compact	AMI Sodium P / Compact	AMU-II pH-Redox
AMI Oxysafe	AMI Solicon4	AMU-II Pharmacon
AMI Oxytrace / Compact	AMI Trides / Compact	AMU-II Powercon
AMI Oxytrace QED / Compact	AMI Turbitrace	AMU Rescon
AMI Pharmacon	AMI Turbitrack	AMU Solicon4
AMI Phosphate-II	AMI Turbiwell	
AMI Phosphate HL		

conform the following UK legislation:

- Electromagnetic Compatibility Regulations 2016
- Electrical Equipment (Safety) Regulations 2016
- Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

Applied harmonized standards:

- BS EN 61326-1:2018 (Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements)
- BS EN 61010-1:2020 (Safety requirements for electrical equipment for measurement, control, and laboratory use)

Hinwil, April 19, 2024

A handwritten signature in blue ink, appearing to read 'T. Burgherr', is written over a light blue horizontal line.

Thomas Burgherr
Corporate Quality Manager
Swan Analytische Instrumente AG